

QUZHOU RONGBO MEDICAL INSTRUMENT CO., LTD.
NO.680 Century Road , Quzhou City, Zhejiang Province, China

TEST REPORT

Product name:	Disposable Medical Face Mask	
Description:	Non-Sterile Type with Ear hook	
Batch size:	46 000 pcs	Control standard: EN 14683:2019
Batch number:	20200320	Date of manufacturing: 20.03. 2020
Validity term:	19.03. 2025	Date of analysis: 21.03. 2020
Certificate number:	QU-202103-1	Date of certificate: 25.03. 2020

CONTROL RESULTS


according to EN 14683:2019 for Type I medical face masks

PARAMETER	STANDARD VALUE	CONTROL RESULT	CONFORM	NOT CONFORM
BACTERIA FILTRATION EFFICIENCY (BFE) %	Type I ≥ 95	≥95	<input checked="" type="checkbox"/>	<input type="checkbox"/>
DIFFERENTIAL PRESSURE (Pa/cm ²)	Type I < 29.4	<29.4	<input checked="" type="checkbox"/>	<input type="checkbox"/>
MICROBIAL CLEANLINESS (cfu/g)	≤ 30	≤ 30	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Storage conditions: In dry place, temperature not higher than 25°C, avoid direct sunlight.

Conclusion: Batch of product was checked according to standard EN 14683:2019 for Type I medical face masks and **conforms** to product specification and requirements set by standard.

Performed by:  (signature)

Confirmed by:  (signature)

SEAL:

